

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090622\
 Data File : VY010352.D
 Acq On : 06 Sep 2022 17:28
 Operator : KP/MD
 Sample : N4500-03
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-25-1.5-2.0

Quant Time: Sep 07 05:55:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 16:41:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	254462	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	422294	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	396961	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	184375	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	126591	48.517	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.040%
35) Dibromofluoromethane	7.710	113	126523	45.005	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.020%
50) Toluene-d8	10.179	98	496374	52.737	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.480%
62) 4-Bromofluorobenzene	12.477	95	170921	44.671	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.340%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090622\
Data File : VY010352.D
Acq On : 06 Sep 2022 17:28
Operator : KP/MD
Sample : N4500-03
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-25-1.5-2.0

Quant Time: Sep 07 05:55:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 16:41:46 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY010352.D
Acq: 06 Sep 2022 17:28




Instrument :
MSVOA_Y
ClientSampleId :
SB-25-1.5-2.0


Tgt Ion:168 Resp: 254462
Ion Ratio Lower Upper
168 100
99 53.1 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 48.517 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VY010352.D
Acq: 06 Sep 2022 17:28

Tgt Ion: 65 Resp: 126591
Ion Ratio Lower Upper
65 100
67 52.7 0.0 99.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.685 min Scan# 1126
Delta R.T. 0.000 min
Lab File: VY010352.D
Acq: 06 Sep 2022 17:28




Instrument :
MSVOA_Y
ClientSampleId :
SB-25-1.5-2.0


Tgt Ion:114 Resp: 422294
Ion Ratio Lower Upper
114 100
63 18.2 0.0 43.2
88 15.0 0.0 31.8



#35
Dibromofluoromethane
Concen: 45.005 ug/l
RT: 7.710 min Scan# 966
Delta R.T. -0.000 min
Lab File: VY010352.D
Acq: 06 Sep 2022 17:28

Tgt Ion:113 Resp: 126523
Ion Ratio Lower Upper
113 100
111 102.7 82.6 124.0
192 20.8 17.6 26.4









#50
Toluene-d8
Concen: 52.737 ug/l
RT: 10.179 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY010352.D
Acq: 06 Sep 2022 17:28




Instrument :
MSVOA_Y
ClientSampleId :
SB-25-1.5-2.0


Tgt Ion: 98 Resp: 496374
Ion Ratio Lower Upper
98 100
100 66.0 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 44.671 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. -0.006 min
Lab File: VY010352.D
Acq: 06 Sep 2022 17:28

Tgt Ion: 95 Resp: 170921
Ion Ratio Lower Upper
95 100
174 87.8 0.0 177.0
176 86.2 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 11
Delta R.T. -0.001 min
Lab File: VY010352.D
Acq: 06 Sep 2022 17:28

Instrument :
MSVOA_Y
ClientSampleId :
SB-25-1.5-2.0

Tgt Ion:117 Resp: 396961
Ion Ratio Lower Upper
117 100
82 51.7 43.0 64.4
119 31.5 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.422 min Scan# 1903
Delta R.T. -0.006 min
Lab File: VY010352.D
Acq: 06 Sep 2022 17:28

Tgt Ion:152 Resp: 184375
Ion Ratio Lower Upper
152 100
115 57.0 28.2 84.7
150 154.9 0.0 347.6

