



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090622\
Data File : VY010340.D
Acq On : 06 Sep 2022 10:33
Operator : KP/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
Title : SW846 8260
Last Update : Fri Sep 02 16:41:46 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1739

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.2	35445	PASS
75	95	30	60	50.9	99213	PASS
95	95	100	100	100.0	194923	PASS
96	95	5	9	6.6	12907	PASS
173	174	0.00	2	0.9	1561	PASS
174	95	50	100	89.3	174101	PASS
175	174	5	9	7.3	12678	PASS
176	174	95	101	96.5	168021	PASS
177	176	5	9	6.8	11412	PASS