

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090721\
 Data File : VY005945.D
 Acq On : 07 Sep 2021 15:38
 Operator : SY/MD
 Sample : M3624-01
 Misc : 5.06G/5ML/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HD-01-090121RE

Quant Time: Sep 07 15:59:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 01 05:32:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	72596	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	113489	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	100663	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	50622	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	35869	49.030	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.060%
35) Dibromofluoromethane	7.728	113	35125	55.111	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.220%
50) Toluene-d8	10.185	98	124874	46.684	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	93.360%
62) 4-Bromofluorobenzene	12.483	95	41417	46.113	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090721\
Data File : VY005945.D
Acq On : 07 Sep 2021 15:38
Operator : SY/MD
Sample : M3624-01
Misc : 5.06G/5ML/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HD-01-090121RE

Quant Time: Sep 07 15:59:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 01 05:32:43 2021
Response via : Initial Calibration

