



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090721\
 Data File : VY005935.D
 Acq On : 07 Sep 2021 09:16
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y0831215.M
 Title : SW846 8260
 Last Update : Wed Sep 01 05:32:43 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.8	23872	PASS
75	95	30	60	48.7	69413	PASS
95	95	100	100	100.0	142469	PASS
96	95	5	9	6.7	9483	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	91.8	130805	PASS
175	174	5	9	7.7	10013	PASS
176	174	95	101	95.9	125387	PASS
177	176	5	9	6.7	8443	PASS