

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090821\
 Data File : VY005957.D
 Acq On : 08 Sep 2021 11:22
 Operator : SY/MD
 Sample : VY0908SBL01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0908SBL01

Quant Time: Sep 09 04:51:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 01 05:32:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	193350	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	324950	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	344295	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	180610	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	103416	53.076	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.160%
35) Dibromofluoromethane	7.728	113	99819	54.698	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.400%
50) Toluene-d8	10.185	98	410145	53.552	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.100%
62) 4-Bromofluorobenzene	12.483	95	154622	60.125	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	120.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090821\
Data File : VY005957.D
Acq On : 08 Sep 2021 11:22
Operator : SY/MD
Sample : VY0908SBL01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0908SBL01

Quant Time: Sep 09 04:51:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 01 05:32:43 2021
Response via : Initial Calibration

