

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090821\
 Data File : VY005961.D
 Acq On : 08 Sep 2021 13:00
 Operator : SY/MD
 Sample : M3658-05RE
 Misc : 4.99G/5ML/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RBR-400027RE

Quant Time: Sep 09 04:53:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 01 05:32:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	116043	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	185700	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	188030	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	93652	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	59794	51.132	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.260%
35) Dibromofluoromethane	7.728	113	57059	54.713	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.420%
50) Toluene-d8	10.185	98	231799	52.960	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.920%
62) 4-Bromofluorobenzene	12.483	95	82605	56.207	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090821\
Data File : VY005961.D
Acq On : 08 Sep 2021 13:00
Operator : SY/MD
Sample : M3658-05RE
Misc : 4.99G/5ML/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RBR-400027RE

Quant Time: Sep 09 04:53:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 01 05:32:43 2021
Response via : Initial Calibration

