

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090821\
 Data File : VY005962.D
 Acq On : 08 Sep 2021 13:23
 Operator : SY/MD
 Sample : M3657-04
 Misc : 5.64G/5ML/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RO-72-11944

Quant Time: Sep 09 04:53:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 01 05:32:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	178742	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	296733	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	316864	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	168168	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	98096	54.461	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.920%	
35) Dibromofluoromethane	7.728	113	91847	55.116	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.240%	
50) Toluene-d8	10.185	98	374401	53.533	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.060%	
62) 4-Bromofluorobenzene	12.483	95	143544	61.125	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 122.240%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090821\
Data File : VY005962.D
Acq On : 08 Sep 2021 13:23
Operator : SY/MD
Sample : M3657-04
Misc : 5.64G/5ML/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RO-72-11944

Quant Time: Sep 09 04:53:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 01 05:32:43 2021
Response via : Initial Calibration

