

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090821\
 Data File : VY005965.D
 Acq On : 08 Sep 2021 14:32
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 09 04:54:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 01 05:32:43 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.446	0.408	8.5	83	0.00
3 P	Chloromethane	0.473	0.412	12.9	84	0.00
4 C	Vinyl Chloride	0.615	0.538	12.5#	85	0.00
5 T	Bromomethane	0.415	0.423	-1.9	100	0.00
6 T	Chloroethane	0.368	0.337	8.4	84	0.00
7 T	Trichlorofluoromethane	0.785	0.806	-2.7	95	0.00
8 T	Diethyl Ether	0.267	0.288	-7.9	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.423	0.478	-13.0	104	0.00
10 T	Methyl Iodide	0.531	0.472	11.1	79	0.00
11 T	Tert butyl alcohol	0.099	0.089	10.1	104	0.00
12 CM	1,1-Dichloroethene	0.429	0.469	-9.3#	101	0.00
13 T	Acrolein	0.021	0.017	19.0	105	0.00
14 T	Allyl chloride	0.722	0.756	-4.7	95	0.00
15 T	Acrylonitrile	0.135	0.144	-6.7	96	0.00
16 T	Acetone	0.128	0.125	2.3	84	0.00
17 T	Carbon Disulfide	1.371	1.433	-4.5	95	0.00
18 T	Methyl Acetate	0.357	0.369	-3.4	97	0.00
19 T	Methyl tert-butyl Ether	1.289	1.427	-10.7	99	0.00
20 T	Methylene Chloride	0.569	0.588	-3.3	103	0.00
21 T	trans-1,2-Dichloroethene	0.479	0.531	-10.9	101	0.00
22 T	Diisopropyl ether	1.467	1.587	-8.2	97	0.00
23 T	Vinyl Acetate	0.890	0.947	-6.4	93	0.00
24 P	1,1-Dichloroethane	0.837	0.920	-9.9	100	0.00
25 T	2-Butanone	0.182	0.187	-2.7	91	0.00
26 T	2,2-Dichloropropane	0.783	0.831	-6.1	97	0.00
27 T	cis-1,2-Dichloroethene	0.538	0.615	-14.3	102	0.00
28 T	Bromochloromethane	0.345	0.335	2.9	82	0.00
29 T	Tetrahydrofuran	0.118	0.109	7.6	83	0.00
30 C	Chloroform	0.858	0.849	1.0#	90	0.00
31 T	Cyclohexane	0.850	0.756	11.1	84	0.00
32 T	1,1,1-Trichloroethane	0.799	0.777	2.8	88	0.00
33 S	1,2-Dichloroethane-d4	0.504	0.465	7.7	83	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
35 S	Dibromofluoromethane	0.281	0.293	-4.3	92	0.00
36 T	1,1-Dichloropropene	0.471	0.459	2.5	86	0.00
37 T	Ethyl Acetate	0.282	0.289	-2.5	91	0.00
38 T	Carbon Tetrachloride	0.510	0.504	1.2	87	0.00
39 T	Methylcyclohexane	0.596	0.588	1.3	87	0.00
40 TM	Benzene	1.330	1.327	0.2	88	0.00
41 T	Methacrylonitrile	0.157	0.149	5.1	82	0.00
42 TM	1,2-Dichloroethane	0.432	0.402	6.9	83	0.00
43 T	Isopropyl Acetate	0.512	0.484	5.5	82	0.00
44 TM	Trichloroethene	0.373	0.381	-2.1	90	0.00
45 C	1,2-Dichloropropane	0.326	0.325	0.3#	89	0.00
46 T	Dibromomethane	0.188	0.190	-1.1	90	0.00
47 T	Bromodichloromethane	0.456	0.471	-3.3	91	0.00
48 T	Methyl methacrylate	0.235	0.221	6.0	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090821\
 Data File : VY005965.D
 Acq On : 08 Sep 2021 14:32
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 09 04:54:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 01 05:32:43 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	92	0.00
50 S	Toluene-d8	1.178	1.189	-0.9	88	0.00
51 T	4-Methyl-2-Pentanone	0.276	0.260	5.8	82	0.00
52 CM	Toluene	0.857	0.863	-0.7#	88	0.00
53 T	t-1,3-Dichloropropene	0.506	0.511	-1.0	88	0.00
54 T	cis-1,3-Dichloropropene	0.555	0.561	-1.1	89	0.00
55 T	1,1,2-Trichloroethane	0.269	0.281	-4.5	92	0.00
56 T	Ethyl methacrylate	0.389	0.402	-3.3	88	0.00
57 T	1,3-Dichloropropane	0.466	0.474	-1.7	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.201	0.197	2.0	87	0.00
59 T	2-Hexanone	0.193	0.183	5.2	80	0.00
60 T	Dibromochloromethane	0.330	0.353	-7.0	93	0.00
61 T	1,2-Dibromoethane	0.256	0.272	-6.3	93	0.00
62 S	4-Bromofluorobenzene	0.396	0.412	-4.0	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.409	0.400	2.2	89	0.00
65 PM	Chlorobenzene	1.003	1.008	-0.5	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.376	0.390	-3.7	93	0.00
67 C	Ethyl Benzene	1.793	1.799	-0.3#	89	0.00
68 T	m/p-Xylenes	0.707	0.707	0.0	89	0.00
69 T	o-Xylene	0.664	0.675	-1.7	90	0.00
70 T	Styrene	1.133	1.152	-1.7	90	0.00
71 P	Bromoform	0.248	0.265	-6.9	94	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	3.333	3.377	-1.3	90	0.00
74 T	N-amyl acetate	0.956	0.899	6.0	81	0.00
75 P	1,1,2,2-Tetrachloroethane	0.634	0.657	-3.6	91	0.00
76 T	1,2,3-Trichloropropane	0.468	0.439	6.2	87	0.00
77 T	Bromobenzene	0.839	0.869	-3.6	93	0.00
78 T	n-propylbenzene	3.983	4.040	-1.4	90	0.00
79 T	2-Chlorotoluene	2.217	2.272	-2.5	91	0.00
80 T	1,3,5-Trimethylbenzene	2.792	2.824	-1.1	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.246	0.241	2.0	84	0.00
82 T	4-Chlorotoluene	2.326	2.335	-0.4	91	0.00
83 T	tert-Butylbenzene	2.441	2.512	-2.9	90	0.00
84 T	1,2,4-Trimethylbenzene	2.770	2.807	-1.3	89	0.00
85 T	sec-Butylbenzene	3.597	3.656	-1.6	89	0.00
86 T	p-Isopropyltoluene	3.117	3.141	-0.8	88	0.00
87 T	1,3-Dichlorobenzene	1.633	1.683	-3.1	91	0.00
88 T	1,4-Dichlorobenzene	1.631	1.654	-1.4	91	0.00
89 T	n-Butylbenzene	2.789	2.815	-0.9	88	0.00
90 T	Hexachloroethane	0.566	0.585	-3.4	93	0.00
91 T	1,2-Dichlorobenzene	1.448	1.471	-1.6	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.119	0.112	5.9	81	0.00
93 T	1,2,4-Trichlorobenzene	0.978	0.973	0.5	88	0.00
94 T	Hexachlorobutadiene	0.630	0.635	-0.8	88	0.00
95 T	Naphthalene	1.878	1.863	0.8	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090821\
Data File : VY005965.D
Acq On : 08 Sep 2021 14:32
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Sep 09 04:54:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 01 05:32:43 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.853	0.851	0.2	86	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6