

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090822\
 Data File : VY010399.D
 Acq On : 08 Sep 2022 16:35
 Operator : KP/MD
 Sample : N4500-07
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-09-1.5-2.0

Quant Time: Sep 09 06:26:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 02 16:41:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	298178	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	513814	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	505222	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	248915	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	139535	45.637	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.280%
35) Dibromofluoromethane	7.716	113	129841	37.959	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	75.920%
50) Toluene-d8	10.179	98	507844	44.345	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.700%
62) 4-Bromofluorobenzene	12.483	95	181855	38.156	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	76.320%
Target Compounds						
20) Methylene Chloride	4.710	84	24706	6.698	ug/l	# 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090822\
Data File : VY010399.D
Acq On : 08 Sep 2022 16:35
Operator : KP/MD
Sample : N4500-07
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-09-1.5-2.0

Quant Time: Sep 09 06:26:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090222S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 02 16:41:46 2022
Response via : Initial Calibration

