

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090921\
 Data File : VY005979.D
 Acq On : 09 Sep 2021 16:51
 Operator : SY/MD
 Sample : M3694-03RE
 Misc : 5.04G/5ML/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-3RE

Quant Time: Sep 10 04:45:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 01 05:32:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	45833	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	68497	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	71095	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	111682	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	21305	46.128	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.260%
35) Dibromofluoromethane	7.722	113	19742	51.321	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.640%
50) Toluene-d8	10.185	98	647473	401.054	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	802.100%#
62) 4-Bromofluorobenzene	12.477	95	417593	770.333	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	1540.660%#
Target Compounds						
					Qvalue	
16) Acetone	3.954	43	2303	19.555	ug/l	93
85) sec-Butylbenzene	13.257	105	233151	29.019	ug/l #	38

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090921\
Data File : VY005979.D
Acq On : 09 Sep 2021 16:51
Operator : SY/MD
Sample : M3694-03RE
Misc : 5.04G/5ML/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-3RE

Quant Time: Sep 10 04:45:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 01 05:32:43 2021
Response via : Initial Calibration

