


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY090925\
Data File : VY023311.D
Acq On : 09 Sep 2025 08:17
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090825S.M
Title : SW846 8260
Last Update : Wed Sep 10 01:44:33 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1744

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	17.5	30445	PASS
75	95	30	60	50.4	87445	PASS
95	95	100	100	100.0	173525	PASS
96	95	5	9	6.5	11243	PASS
173	174	0.00	2	1.1	1721	PASS
174	95	50	100	90.6	157269	PASS
175	174	5	9	7.3	11539	PASS
176	174	95	101	97.0	152504	PASS
177	176	5	9	6.4	9719	PASS