

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091021\
 Data File : VY005983.D
 Acq On : 10 Sep 2021 15:15
 Operator : SY/MD
 Sample : VY0910SBL01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0910SBL01

Quant Time: Sep 11 01:20:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 01 05:32:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	221675	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	372926	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	397513	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	206595	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	117077	52.410	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.820%
35) Dibromofluoromethane	7.722	113	116974	55.852	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.700%
50) Toluene-d8	10.179	98	470782	53.561	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.120%
62) 4-Bromofluorobenzene	12.483	95	179880	60.948	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	121.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091021\
Data File : VY005983.D
Acq On : 10 Sep 2021 15:15
Operator : SY/MD
Sample : VY0910SBL01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0910SBL01

Quant Time: Sep 11 01:20:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 01 05:32:43 2021
Response via : Initial Calibration

