

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091021\
 Data File : VY005986.D
 Acq On : 10 Sep 2021 16:30
 Operator : SY/MD
 Sample : M3684-16
 Misc : 5.11G/5ML/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 A007015

Quant Time: Sep 11 01:22:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 01 05:32:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	209549	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	348204	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	308740	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	96092	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	116698	55.263	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.520%
35) Dibromofluoromethane	7.728	113	114130	58.364	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	116.720%
50) Toluene-d8	10.185	98	431609	52.591	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.180%
62) 4-Bromofluorobenzene	12.483	95	110910	40.247	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091021\
Data File : VY005986.D
Acq On : 10 Sep 2021 16:30
Operator : SY/MD
Sample : M3684-16
Misc : 5.11G/5ML/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
A007015

Quant Time: Sep 11 01:22:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 01 05:32:43 2021
Response via : Initial Calibration

