

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091021\
 Data File : VY005987.D
 Acq On : 10 Sep 2021 16:53
 Operator : SY/MD
 Sample : M3684-17
 Misc : 5.38G/5ML/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 A007145

Quant Time: Sep 11 01:22:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 01 05:32:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	221880	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	359076	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	345097	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	163503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	90478	40.465	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	80.940%
35) Dibromofluoromethane	7.722	113	103462	51.306	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.620%
50) Toluene-d8	10.185	98	440508	52.050	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.100%
62) 4-Bromofluorobenzene	12.483	95	148261	52.172	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091021\
Data File : VY005987.D
Acq On : 10 Sep 2021 16:53
Operator : SY/MD
Sample : M3684-17
Misc : 5.38G/5ML/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
A007145

Quant Time: Sep 11 01:22:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 01 05:32:43 2021
Response via : Initial Calibration

