

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091021\
 Data File : VY005990.D
 Acq On : 10 Sep 2021 18:03
 Operator : SY/MD
 Sample : M3684-20
 Misc : 5.61G/5ML/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 A006015

Quant Time: Sep 11 01:23:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 01 05:32:43 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	192162	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	322854	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	328532	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	143751	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	105101	54.275	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	108.540%
35) Dibromofluoromethane	7.722	113	101880	56.190	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.380%
50) Toluene-d8	10.185	98	404427	53.148	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.300%
62) 4-Bromofluorobenzene	12.483	95	135724	53.119	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091021\
Data File : VY005990.D
Acq On : 10 Sep 2021 18:03
Operator : SY/MD
Sample : M3684-20
Misc : 5.61G/5ML/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
A006015

Quant Time: Sep 11 01:23:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y083121S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 01 05:32:43 2021
Response via : Initial Calibration

