

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091024\
 Data File : VY019484.D
 Acq On : 10 Sep 2024 13:50
 Operator : SY/MD
 Sample : P3906-03
 Misc : 4.84g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-701-VOC-01

Quant Time: Sep 11 01:49:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	130086	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	216948	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	163443	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	58161	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	56582	46.973	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.940%
35) Dibromofluoromethane	7.634	113	62769	51.029	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.060%
50) Toluene-d8	10.109	98	245658	53.994	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	107.980%
62) 4-Bromofluorobenzene	12.408	95	71437	40.835	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	81.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091024\
Data File : VY019484.D
Acq On : 10 Sep 2024 13:50
Operator : SY/MD
Sample : P3906-03
Misc : 4.84g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SU-701-VOC-01

Quant Time: Sep 11 01:49:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

