

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091024\
 Data File : VY019486.D
 Acq On : 10 Sep 2024 14:37
 Operator : SY/MD
 Sample : P3892-01RE
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ARS20-005RE

Quant Time: Sep 11 01:49:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	49057	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	77797	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	49907	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	13136	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	18227	40.125	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	80.240%
35) Dibromofluoromethane	7.634	113	21280	48.244	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.480%
50) Toluene-d8	10.103	98	82679	50.676	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.360%
62) 4-Bromofluorobenzene	12.402	95	20434	32.573	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	65.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091024\
Data File : VY019486.D
Acq On : 10 Sep 2024 14:37
Operator : SY/MD
Sample : P3892-01RE
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ARS20-005RE

Quant Time: Sep 11 01:49:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

