

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091123\
 Data File : VY015501.D
 Acq On : 11 Sep 2023 20:49
 Operator : SY/MD
 Sample : 04351-01
 Misc : 4.44g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CPT101-02-(2-2.5)

Quant Time: Sep 12 01:23:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	313024	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	514655	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	465973	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	223685	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	196562	58.578	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.160%
35) Dibromofluoromethane	7.716	113	161332	54.425	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.840%
50) Toluene-d8	10.179	98	604852	56.345	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	112.680%
62) 4-Bromofluorobenzene	12.477	95	207889	52.166	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	104.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091123\
Data File : VY015501.D
Acq On : 11 Sep 2023 20:49
Operator : SY/MD
Sample : 04351-01
Misc : 4.44g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CPT101-02-(2-2.5)

Quant Time: Sep 12 01:23:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 31 20:01:47 2023
Response via : Initial Calibration

