

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091223\
 Data File : VY015509.D
 Acq On : 12 Sep 2023 15:47
 Operator : SY/MD
 Sample : 04393-01
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-1

Quant Time: Sep 13 05:54:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	295502	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	495823	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	453720	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	216403	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	188307	59.445	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.900%
35) Dibromofluoromethane	7.710	113	154549	54.117	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.240%
50) Toluene-d8	10.179	98	578062	55.894	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	111.780%
62) 4-Bromofluorobenzene	12.477	95	202010	52.616	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	105.240%
Target Compounds						
16) Acetone	3.942	43	21284	16.097	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091223\
Data File : VY015509.D
Acq On : 12 Sep 2023 15:47
Operator : SY/MD
Sample : 04393-01
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-1

Quant Time: Sep 13 05:54:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 31 20:01:47 2023
Response via : Initial Calibration

