

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091223\
 Data File : VY015510.D
 Acq On : 12 Sep 2023 16:11
 Operator : SY/MD
 Sample : 04395-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-01-91123

Quant Time: Sep 13 05:54:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.782	168	181696	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	300887	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	262187	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	119160	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	111626	57.310	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 114.620%	
35) Dibromofluoromethane	7.709	113	93873	54.166	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.340%	
50) Toluene-d8	10.178	98	346464	55.205	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 110.400%	
62) 4-Bromofluorobenzene	12.477	95	113330	48.642	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 97.280%	
Target Compounds						
16) Acetone	3.942	43	15477	19.037	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091223\
Data File : VY015510.D
Acq On : 12 Sep 2023 16:11
Operator : SY/MD
Sample : 04395-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-01-91123

Quant Time: Sep 13 05:54:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 31 20:01:47 2023
Response via : Initial Calibration

