

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091223\
 Data File : VY015524.D
 Acq On : 12 Sep 2023 21:45
 Operator : SY/MD
 Sample : 04397-05
 Misc : 4.34g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CPT103-07(8-8.5)

Quant Time: Sep 13 05:58:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	324349	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	536600	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	478907	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	216706	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	200579	57.688	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.380%
35) Dibromofluoromethane	7.716	113	165929	53.686	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.380%
50) Toluene-d8	10.179	98	630487	56.331	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	112.660%
62) 4-Bromofluorobenzene	12.477	95	205447	49.445	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	98.880%
Target Compounds						
					Qvalue	
16) Acetone	3.948	43	18457	12.718	ug/l	95
17) Carbon Disulfide	4.186	76	63338	6.571	ug/l	100
25) 2-Butanone	6.984	43	5770	2.803	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091223\
Data File : VY015524.D
Acq On : 12 Sep 2023 21:45
Operator : SY/MD
Sample : 04397-05
Misc : 4.34g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CPT103-07(8-8.5)

Quant Time: Sep 13 05:58:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 31 20:01:47 2023
Response via : Initial Calibration

