

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091223\
 Data File : VY015525.D
 Acq On : 12 Sep 2023 22:08
 Operator : SY/MD
 Sample : 04397-07
 Misc : 3.86g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B103-03(10-10.5)

Quant Time: Sep 13 05:58:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	322329	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	534845	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	483470	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	228723	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	204357	59.143	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 118.280%	
35) Dibromofluoromethane	7.716	113	165868	53.843	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.680%	
50) Toluene-d8	10.179	98	629357	56.414	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 112.820%	
62) 4-Bromofluorobenzene	12.483	95	214751	51.853	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 103.700%	
Target Compounds						
17) Carbon Disulfide	4.192	76	45990	4.801	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091223\
Data File : VY015525.D
Acq On : 12 Sep 2023 22:08
Operator : SY/MD
Sample : 04397-07
Misc : 3.86g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B103-03(10-10.5)

Quant Time: Sep 13 05:58:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 31 20:01:47 2023
Response via : Initial Calibration

