

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091223\  
 Data File : VY015520.D  
 Acq On : 12 Sep 2023 20:09  
 Operator : SY/MD  
 Sample : 04394-09  
 Misc : 6.15g/5.0mL/MSVOA\_Y/SOIL  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 B103-08(3-.35)

Quant Time: Sep 13 05:57:15 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y082923S.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Aug 31 20:01:47 2023  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 7.782  | 168  | 321226   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 8.685  | 114  | 532244   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 11.489 | 117  | 485829   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 13.422 | 152  | 230992   | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 33) 1,2-Dichloroethane-d4   | 8.136  | 65    | 203176   | 59.003   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 50 - 163 | Recovery | =    | 118.000% |
| 35) Dibromofluoromethane    | 7.715  | 113   | 167449   | 54.622   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 54 - 147 | Recovery | =    | 109.240% |
| 50) Toluene-d8              | 10.178 | 98    | 622112   | 56.037   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 134 | Recovery | =    | 112.080% |
| 62) 4-Bromofluorobenzene    | 12.477 | 95    | 214112   | 51.952   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 30 - 143 | Recovery | =    | 103.900% |


| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY091223\  
Data File : VY015520.D  
Acq On : 12 Sep 2023 20:09  
Operator : SY/MD  
Sample : 04394-09  
Misc : 6.15g/5.0mL/MSVOA\_Y/SOIL  
ALS Vial : 19 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
B103-08(3-.35)

Quant Time: Sep 13 05:57:15 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y082923S.M  
Quant Title : SW846 8260  
QLast Update : Thu Aug 31 20:01:47 2023  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 7.782 min Scan# 91  
Delta R.T. -0.001 min  
Lab File: VY015520.D  
Acq: 12 Sep 2023 20:09




Instrument :  
MSVOA\_Y  
ClientSampleId :  
B103-08(3-.35)


Tgt Ion:168 Resp: 321226  
Ion Ratio Lower Upper  
168 100  
99 56.7 43.4 65.0



#33  
1,2-Dichloroethane-d4  
Concen: 59.003 ug/l  
RT: 8.136 min Scan# 1036  
Delta R.T. -0.000 min  
Lab File: VY015520.D  
Acq: 12 Sep 2023 20:09

Tgt Ion: 65 Resp: 203176  
Ion Ratio Lower Upper  
65 100  
67 51.9 0.0 101.8









#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 8.685 min Scan# 1126  
Delta R.T. -0.000 min  
Lab File: VY015520.D  
Acq: 12 Sep 2023 20:09




Instrument :  
MSVOA\_Y  
ClientSampleId :  
B103-08(3-35)


Tgt Ion:114 Resp: 532244  
Ion Ratio Lower Upper  
114 100  
63 19.5 0.0 42.4  
88 15.0 0.0 30.6



#35  
Dibromofluoromethane  
Concen: 54.622 ug/l  
RT: 7.715 min Scan# 967  
Delta R.T. -0.000 min  
Lab File: VY015520.D  
Acq: 12 Sep 2023 20:09

Tgt Ion:113 Resp: 167449  
Ion Ratio Lower Upper  
113 100  
111 102.6 81.3 121.9  
192 19.5 16.9 25.3









#50  
Toluene-d8  
Concen: 56.037 ug/l  
RT: 10.178 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VY015520.D  
Acq: 12 Sep 2023 20:09




Instrument :  
MSVOA\_Y  
ClientSampleId :  
B103-08(3-.35)


Tgt Ion: 98 Resp: 622112  
Ion Ratio Lower Upper  
98 100  
100 66.3 50.1 75.1



#62  
4-Bromofluorobenzene  
Concen: 51.952 ug/l  
RT: 12.477 min Scan# 1748  
Delta R.T. -0.000 min  
Lab File: VY015520.D  
Acq: 12 Sep 2023 20:09

Tgt Ion: 95 Resp: 214112  
Ion Ratio Lower Upper  
95 100  
174 86.0 0.0 178.2  
176 82.0 0.0 173.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.489 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY015520.D

Acq: 12 Sep 2023 20:09

Instrument :

MSVOA\_Y

ClientSampleId :

B103-08(3-35)





Tgt Ion:117 Resp: 485829

Ion Ratio Lower Upper

117 100

82 52.8 43.4 65.0

119 31.9 26.3 39.5



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.422 min Scan# 1903

Delta R.T. -0.000 min

Lab File: VY015520.D

Acq: 12 Sep 2023 20:09

Tgt Ion:152 Resp: 230992

Ion Ratio Lower Upper

152 100

115 57.0 28.8 86.5

150 155.6 0.0 348.8

