



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091223\
Data File : VY015503.D
Acq On : 12 Sep 2023 11:21
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Title : SW846 8260
Last Update : Thu Aug 31 20:01:47 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.2	20839	PASS
75	95	30	60	52.9	57435	PASS
95	95	100	100	100.0	108632	PASS
96	95	5	9	6.6	7155	PASS
173	174	0.00	2	1.0	952	PASS
174	95	50	100	88.1	95677	PASS
175	174	5	9	7.3	6990	PASS
176	174	95	101	97.4	93163	PASS
177	176	5	9	6.8	6331	PASS