

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091322\
 Data File : VY010463.D
 Acq On : 13 Sep 2022 16:55
 Operator : KP/MD
 Sample : N4509-04RE
 Misc : 5.59g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S1-SO_115D-20220902RE

Quant Time: Sep 14 05:25:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 09 17:08:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	209519	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	353715	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	321878	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	139225	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	115952	54.948	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	109.900%
35) Dibromofluoromethane	7.709	113	109294	43.507	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.020%
50) Toluene-d8	10.179	98	422388	51.880	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.760%
62) 4-Bromofluorobenzene	12.477	95	138501	41.923	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	83.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091322\
Data File : VY010463.D
Acq On : 13 Sep 2022 16:55
Operator : KP/MD
Sample : N4509-04RE
Misc : 5.59g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S1-SO_115D-20220902RE

Quant Time: Sep 14 05:25:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 09 17:08:30 2022
Response via : Initial Calibration

