

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091322\
 Data File : VY010465.D
 Acq On : 13 Sep 2022 17:42
 Operator : KP/MD
 Sample : N4509-06
 Misc : 5.37g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S1-SO_PW001-20220902

Quant Time: Sep 14 05:26:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 09 17:08:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	211944	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	358520	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	323073	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	138117	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	114824	53.791	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.580%
35) Dibromofluoromethane	7.710	113	109274	42.916	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.840%
50) Toluene-d8	10.173	98	423239	51.288	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.580%
62) 4-Bromofluorobenzene	12.477	95	135148	40.098	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091322\
Data File : VY010465.D
Acq On : 13 Sep 2022 17:42
Operator : KP/MD
Sample : N4509-06
Misc : 5.37g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S1-SO_PW001-20220902

Quant Time: Sep 14 05:26:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 09 17:08:30 2022
Response via : Initial Calibration

