

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091322\
 Data File : VY010469.D
 Acq On : 13 Sep 2022 19:16
 Operator : KP/MD
 Sample : N4466-02
 Misc : 5.56g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SASP2-T-5-(17-23)

Quant Time: Sep 14 05:27:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 09 17:08:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	187189	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	321024	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	306237	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	146244	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	111483	59.133	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.260%
35) Dibromofluoromethane	7.710	113	101827	44.662	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.320%
50) Toluene-d8	10.173	98	384464	52.031	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.060%
62) 4-Bromofluorobenzene	12.477	95	134398	45.310	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	90.620%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091322\
Data File : VY010469.D
Acq On : 13 Sep 2022 19:16
Operator : KP/MD
Sample : N4466-02
Misc : 5.56g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SASP2-T-5-(17-23)

Quant Time: Sep 14 05:27:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 09 17:08:30 2022
Response via : Initial Calibration

