


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091322\
 Data File : VY010451.D
 Acq On : 13 Sep 2022 09:57
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
 Title : SW846 8260
 Last Update : Fri Sep 09 17:08:30 2022



AutoFind: Scans 1747, 1748, 1749; Background Corrected with Scan 1739

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.9	17773	PASS
75	95	30	60	51.5	48387	PASS
95	95	100	100	100.0	93909	PASS
96	95	5	9	6.9	6521	PASS
173	174	0.00	2	1.1	858	PASS
174	95	50	100	83.9	78821	PASS
175	174	5	9	7.7	6096	PASS
176	174	95	101	95.9	75560	PASS
177	176	5	9	7.0	5271	PASS