

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091421\
 Data File : VY006012.D
 Acq On : 14 Sep 2021 16:54
 Operator : SY/MD
 Sample : M3727-02
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOC

Quant Time: Sep 15 05:15:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 15 05:01:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	102487	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	191657	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	207327	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	104541	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	81788	57.451	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.900%
35) Dibromofluoromethane	7.728	113	59236	49.761	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.520%
50) Toluene-d8	10.185	98	254343	50.131	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.260%
62) 4-Bromofluorobenzene	12.483	95	93053	54.020	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.040%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091421\
Data File : VY006012.D
Acq On : 14 Sep 2021 16:54
Operator : SY/MD
Sample : M3727-02
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOC

Quant Time: Sep 15 05:15:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 15 05:01:47 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.801 min Scan# 981
Delta R.T. 0.006 min
Lab File: VY006012.D
Acq: 14 Sep 2021 16:54




Instrument :
MSVOA_Y
ClientSampleId :
VOC


Tgt Ion:168 Resp: 102487
Ion Ratio Lower Upper
168 100
99 54.9 46.9 70.3



#33
1,2-Dichloroethane-d4
Concen: 57.451 ug/l
RT: 8.149 min Scan# 1038
Delta R.T. -0.000 min
Lab File: VY006012.D
Acq: 14 Sep 2021 16:54

Tgt Ion: 65 Resp: 81788
Ion Ratio Lower Upper
65 100
67 50.0 0.0 99.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.697 min Scan# 1128
Delta R.T. -0.000 min
Lab File: VY006012.D
Acq: 14 Sep 2021 16:54




Instrument :
MSVOA_Y
ClientSampleId :
VOC


Tgt Ion:	114	Resp:	191657
Ion	Ratio	Lower	Upper
114	100		
63	22.0	0.0	43.2
88	15.0	0.0	31.8



#35
Dibromofluoromethane
Concen: 49.761 ug/l
RT: 7.728 min Scan# 969
Delta R.T. -0.000 min
Lab File: VY006012.D
Acq: 14 Sep 2021 16:54

Tgt Ion:	113	Resp:	59236
Ion	Ratio	Lower	Upper
113	100		
111	103.3	82.6	124.0
192	21.9	17.6	26.4









#50
Toluene-d8
Concen: 50.131 ug/l
RT: 10.185 min Scan# 1372
Delta R.T. -0.000 min
Lab File: VY006012.D
Acq: 14 Sep 2021 16:54




Instrument :
MSVOA_Y
ClientSampled :
VOC


Tgt Ion: 98 Resp: 254343
Ion Ratio Lower Upper
98 100
100 62.5 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 54.020 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VY006012.D
Acq: 14 Sep 2021 16:54

Tgt Ion: 95 Resp: 93053
Ion Ratio Lower Upper
95 100
174 91.4 0.0 177.0
176 88.3 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.496 min Scan# 1587
Delta R.T. -0.000 min
Lab File: VY006012.D
Acq: 14 Sep 2021 16:54

Instrument :
MSVOA_Y
ClientSampleId :
VOC

Tgt Ion:117 Resp: 207327
Ion Ratio Lower Upper
117 100
82 53.0 43.0 64.4
119 31.0 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. -0.000 min
Lab File: VY006012.D
Acq: 14 Sep 2021 16:54

Tgt Ion:152 Resp: 104541
Ion Ratio Lower Upper
152 100
115 57.6 28.2 84.7
150 156.0 0.0 347.6

