

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091422\
 Data File : VY010476.D
 Acq On : 14 Sep 2022 11:12
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 14 15:05:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 09 17:08:30 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.383	0.329	14.1	78	0.00
3 P	Chloromethane	0.562	0.526	6.4	82	0.00
4 C	Vinyl Chloride	0.693	0.651	6.1#	80	0.00
5 T	Bromomethane	0.486	0.441	9.3	80	0.00
6 T	Chloroethane	0.464	0.450	3.0	82	0.00
7 T	Trichlorofluoromethane	0.878	0.842	4.1	81	0.00
8 T	Diethyl Ether	0.284	0.280	1.4	84	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.514	0.504	1.9	82	0.00
10 T	Methyl Iodide	0.636	0.613	3.6	76	0.00
11 T	Tert butyl alcohol	0.068	0.051	25.0#	84	0.00
12 CM	1,1-Dichloroethene	0.482	0.460	4.6#	80	0.00
13 T	Acrolein	0.048	0.045	6.3	78	0.00
14 T	Allyl chloride	0.694	0.705	-1.6	84	0.00
15 T	Acrylonitrile	0.138	0.149	-8.0	91	0.00
16 T	Acetone	0.116	0.133	-14.7	98	0.00
17 T	Carbon Disulfide	1.304	1.166	10.6	76	0.00
18 T	Methyl Acetate	0.356	0.354	0.6	89	0.00
19 T	Methyl tert-butyl Ether	1.332	1.371	-2.9	85	0.00
20 T	Methylene Chloride	0.705	0.607	13.9	88	0.00
21 T	trans-1,2-Dichloroethene	0.547	0.519	5.1	80	0.00
22 T	Diisopropyl ether	1.550	1.636	-5.5	88	0.00
23 T	Vinyl Acetate	0.950	1.028	-8.2	88	0.00
24 P	1,1-Dichloroethane	0.938	0.959	-2.2	84	0.00
25 T	2-Butanone	0.184	0.199	-8.2	93	0.00
26 T	2,2-Dichloropropane	0.896	0.909	-1.5	87	0.00
27 T	cis-1,2-Dichloroethene	0.636	0.628	1.3	82	0.00
28 T	Bromochloromethane	0.301	0.325	-8.0	83	0.00
29 T	Tetrahydrofuran	0.115	0.124	-7.8	92	0.00
30 C	Chloroform	0.995	1.026	-3.1#	85	0.00
31 T	Cyclohexane	0.851	0.788	7.4	82	0.00
32 T	1,1,1-Trichloroethane	0.901	0.915	-1.6	83	0.00
33 S	1,2-Dichloroethane-d4	0.504	0.517	-2.6	85	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.355	0.321	9.6	84	0.00
36 T	1,1-Dichloropropene	0.486	0.471	3.1	83	0.00
37 T	Ethyl Acetate	0.256	0.274	-7.0	93	0.00
38 T	Carbon Tetrachloride	0.541	0.524	3.1	81	0.00
39 T	Methylcyclohexane	0.606	0.569	6.1	80	0.00
40 TM	Benzene	1.399	1.369	2.1	83	0.00
41 T	Methacrylonitrile	0.146	0.158	-8.2	90	0.01
42 TM	1,2-Dichloroethane	0.416	0.421	-1.2	86	0.00
43 T	Isopropyl Acetate	0.468	0.489	-4.5	89	0.00
44 TM	Trichloroethene	0.412	0.385	6.6	79	0.00
45 C	1,2-Dichloropropane	0.341	0.347	-1.8#	86	0.00
46 T	Dibromomethane	0.206	0.204	1.0	84	0.00
47 T	Bromodichloromethane	0.489	0.500	-2.2	85	0.00
48 T	Methyl methacrylate	0.210	0.221	-5.2	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091422\
 Data File : VY010476.D
 Acq On : 14 Sep 2022 11:12
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 14 15:05:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 09 17:08:30 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	83	0.00
50 S	Toluene-d8	1.151	1.176	-2.2	82	0.00
51 T	4-Methyl-2-Pentanone	0.257	0.276	-7.4	91	0.00
52 CM	Toluene	0.916	0.893	2.5#	83	0.00
53 T	t-1,3-Dichloropropene	0.503	0.507	-0.8	85	0.00
54 T	cis-1,3-Dichloropropene	0.567	0.572	-0.9	84	0.00
55 T	1,1,2-Trichloroethane	0.293	0.294	-0.3	85	0.00
56 T	Ethyl methacrylate	0.385	0.394	-2.3	86	0.00
57 T	1,3-Dichloropropane	0.487	0.499	-2.5	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.131	0.102	22.1	90	0.00
59 T	2-Hexanone	0.175	0.191	-9.1	92	0.00
60 T	Dibromochloromethane	0.358	0.359	-0.3	83	0.00
61 T	1,2-Dibromoethane	0.284	0.277	2.5	83	0.00
62 S	4-Bromofluorobenzene	0.528	0.415	21.4	82	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	85	0.00
64 T	Tetrachloroethene	0.451	0.418	7.3	79	0.00
65 PM	Chlorobenzene	1.118	1.083	3.1	83	0.00
66 T	1,1,1,2-Tetrachloroethane	0.414	0.415	-0.2	83	0.00
67 C	Ethyl Benzene	1.979	1.942	1.9#	83	0.00
68 T	m/p-Xylenes	0.777	0.756	2.7	82	0.00
69 T	o-Xylene	0.742	0.728	1.9	83	0.00
70 T	Styrene	1.256	1.241	1.2	83	0.00
71 P	Bromoform	0.259	0.257	0.8	83	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
73 T	Isopropylbenzene	3.901	3.825	1.9	82	0.00
74 T	N-amyl acetate	0.910	0.957	-5.2	86	0.00
75 P	1,1,2,2-Tetrachloroethane	0.730	0.752	-3.0	88	0.00
76 T	1,2,3-Trichloropropane	0.552	0.597	-8.2	93	0.00
77 T	Bromobenzene	0.924	0.897	2.9	82	0.00
78 T	n-propylbenzene	4.620	4.585	0.8	83	0.00
79 T	2-Chlorotoluene	2.595	2.574	0.8	83	0.00
80 T	1,3,5-Trimethylbenzene	3.292	3.242	1.5	83	0.00
81 T	trans-1,4-Dichloro-2-butene	0.249	0.259	-4.0	87	0.00
82 T	4-Chlorotoluene	2.700	2.657	1.6	84	0.00
83 T	tert-Butylbenzene	2.877	2.823	1.9	82	0.00
84 T	1,2,4-Trimethylbenzene	3.221	3.204	0.5	84	0.00
85 T	sec-Butylbenzene	4.260	4.224	0.8	83	0.00
86 T	p-Isopropyltoluene	3.603	3.554	1.4	82	0.00
87 T	1,3-Dichlorobenzene	1.856	1.778	4.2	81	0.00
88 T	1,4-Dichlorobenzene	1.838	1.768	3.8	82	0.00
89 T	n-Butylbenzene	3.271	3.272	-0.0	83	0.00
90 T	Hexachloroethane	0.677	0.660	2.5	83	0.00
91 T	1,2-Dichlorobenzene	1.644	1.581	3.8	81	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.123	0.125	-1.6	90	0.00
93 T	1,2,4-Trichlorobenzene	1.021	0.979	4.1	81	0.00
94 T	Hexachlorobutadiene	0.616	0.566	8.1	79	0.00
95 T	Naphthalene	2.014	2.013	0.0	85	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091422\
Data File : VY010476.D
Acq On : 14 Sep 2022 11:12
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Sep 14 15:05:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090922S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 09 17:08:30 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.891	0.833	6.5	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6