

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091423\
 Data File : VY015538.D
 Acq On : 14 Sep 2023 17:03
 Operator : SY/MD
 Sample : 04394-03
 Misc : 5.96g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B103-07(3.5-4)

Quant Time: Sep 15 04:03:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	407045	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	678583	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	625502	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	287920	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	249803	57.249	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	114.500%	
35) Dibromofluoromethane	7.716	113	215575	55.155	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	110.320%	
50) Toluene-d8	10.179	98	806494	56.979	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	113.960%	
62) 4-Bromofluorobenzene	12.477	95	290343	55.256	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	110.520%	
Target Compounds						Qvalue
84) 1,2,4-Trimethylbenzene	13.117	105	29180	1.625	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091423\
Data File : VY015538.D
Acq On : 14 Sep 2023 17:03
Operator : SY/MD
Sample : 04394-03
Misc : 5.96g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B103-07(3.5-4)

Quant Time: Sep 15 04:03:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 31 20:01:47 2023
Response via : Initial Calibration

