

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091423\
 Data File : VY015552.D
 Acq On : 14 Sep 2023 22:35
 Operator : SY/MD
 Sample : 04409-15
 Misc : 5.43g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-5(0-6)GRAB

Quant Time: Sep 15 04:08:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	361625	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	601292	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	550076	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	260734	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	220089	56.774	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.540%
35) Dibromofluoromethane	7.716	113	189635	54.755	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.520%
50) Toluene-d8	10.179	98	708255	56.471	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	112.940%
62) 4-Bromofluorobenzene	12.483	95	248056	53.276	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	106.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091423\
Data File : VY015552.D
Acq On : 14 Sep 2023 22:35
Operator : SY/MD
Sample : 04409-15
Misc : 5.43g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-5(0-6)GRAB

Quant Time: Sep 15 04:08:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 31 20:01:47 2023
Response via : Initial Calibration

