

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091523\
 Data File : VY015563.D
 Acq On : 15 Sep 2023 12:43
 Operator : SY/MD
 Sample : 04409-06
 Misc : 6.40g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-2(0-2)GRAB

Quant Time: Sep 15 15:36:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	306955	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	516750	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	462992	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	203481	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	190797	57.984	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	115.960%	
35) Dibromofluoromethane	7.716	113	162510	54.600	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	109.200%	
50) Toluene-d8	10.179	98	601792	55.832	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	111.660%	
62) 4-Bromofluorobenzene	12.483	95	201090	50.255	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	100.520%	
Target Compounds						
29) Tetrahydrofuran	7.356	42	1350	1.104	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091523\
Data File : VY015563.D
Acq On : 15 Sep 2023 12:43
Operator : SY/MD
Sample : 04409-06
Misc : 6.40g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-2(0-2)GRAB

Quant Time: Sep 15 15:36:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 31 20:01:47 2023
Response via : Initial Calibration

