

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091523\
 Data File : VY015578.D
 Acq On : 15 Sep 2023 18:36
 Operator : SY/MD
 Sample : 04435-04
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-1

Quant Time: Sep 15 23:00:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	94451	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	149485	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	117806	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	46828	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	48116	47.522	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.040%
35) Dibromofluoromethane	7.716	113	45221	52.521	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.040%
50) Toluene-d8	10.179	98	169684	54.421	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	108.840%
62) 4-Bromofluorobenzene	12.477	95	45167	39.021	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	78.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091523\
Data File : VY015578.D
Acq On : 15 Sep 2023 18:36
Operator : SY/MD
Sample : 04435-04
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-1

Quant Time: Sep 15 23:00:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 31 20:01:47 2023
Response via : Initial Calibration

