

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091523\
 Data File : VY015596.D
 Acq On : 16 Sep 2023 02:46
 Operator : SY/MD
 Sample : 04419-08
 Misc : 6.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DB-05(23-25)

Quant Time: Sep 16 05:28:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	258160	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.703	114	437324	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.502	117	395908	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.434	152	182082	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	157464	56.899	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	= 113.800%	
35) Dibromofluoromethane	7.728	113	138338	54.920	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.840%	
50) Toluene-d8	10.191	98	511708	56.097	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	= 112.200%	
62) 4-Bromofluorobenzene	12.489	95	170350	50.305	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	= 100.600%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091523\
Data File : VY015596.D
Acq On : 16 Sep 2023 02:46
Operator : SY/MD
Sample : 04419-08
Misc : 6.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DB-05(23-25)

Quant Time: Sep 16 05:28:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 31 20:01:47 2023
Response via : Initial Calibration

