

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091523\
 Data File : VY015599.D
 Acq On : 16 Sep 2023 03:57
 Operator : SY/MD
 Sample : 04419-11
 Misc : 4.74g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DB-06(23-25)

Quant Time: Sep 18 02:53:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	211071	50.000	ug/l	0.02
34) 1,4-Difluorobenzene	8.703	114	360776	50.000	ug/l	0.02
63) Chlorobenzene-d5	11.496	117	331971	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	155443	50.000	ug/l	0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	137519	60.778	ug/l	0.02
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.560%
35) Dibromofluoromethane	7.728	113	113803	54.766	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.540%
50) Toluene-d8	10.191	98	424084	56.355	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	112.720%
62) 4-Bromofluorobenzene	12.489	95	145215	51.981	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	103.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091523\
Data File : VY015599.D
Acq On : 16 Sep 2023 03:57
Operator : SY/MD
Sample : 04419-11
Misc : 4.74g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DB-06(23-25)

Quant Time: Sep 18 02:53:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 31 20:01:47 2023
Response via : Initial Calibration

