

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091523\
 Data File : VY015568.D
 Acq On : 15 Sep 2023 14:41
 Operator : SY/MD
 Sample : 04409-18
 Misc : 6.28g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-6(0-6)GRAB

Quant Time: Sep 15 15:38:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	297620	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	498242	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	454418	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	211315	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	189691	59.456	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.920%
35) Dibromofluoromethane	7.710	113	157977	55.049	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.100%
50) Toluene-d8	10.179	98	585547	56.343	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	112.680%
62) 4-Bromofluorobenzene	12.483	95	201007	52.100	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	104.200%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091523\
Data File : VY015568.D
Acq On : 15 Sep 2023 14:41
Operator : SY/MD
Sample : 04409-18
Misc : 6.28g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-6(0-6)GRAB

Quant Time: Sep 15 15:38:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 31 20:01:47 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY015568.D
Acq: 15 Sep 2023 14:41




Instrument :
MSVOA_Y
ClientSampleId :
WC-6(0-6)GRAB

Tgt Ion:168 Resp: 297620
Ion Ratio Lower Upper
168 100
99 52.1 43.4 65.0



#33
1,2-Dichloroethane-d4
Concen: 59.456 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VY015568.D
Acq: 15 Sep 2023 14:41

Tgt Ion: 65 Resp: 189691
Ion Ratio Lower Upper
65 100
67 52.8 0.0 101.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VY015568.D

Acq: 15 Sep 2023 14:41

Instrument :

MSVOA_Y

ClientSampleId :

WC-6(0-6)GRAB





Tgt Ion:114 Resp: 498242

Ion Ratio Lower Upper

114 100

63 19.5 0.0 42.4

88 15.6 0.0 30.6



#35

Dibromofluoromethane

Concen: 55.049 ug/l

RT: 7.710 min Scan# 966

Delta R.T. -0.006 min

Lab File: VY015568.D

Acq: 15 Sep 2023 14:41




Tgt Ion:113 Resp: 157977

Ion Ratio Lower Upper

113 100

111 103.3 81.3 121.9

192 19.5 16.9 25.3








#50
Toluene-d8
Concen: 56.343 ug/l
RT: 10.179 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY015568.D
Acq: 15 Sep 2023 14:41




Instrument :
MSVOA_Y
ClientSampleId :
WC-6(0-6)GRAB


Tgt Ion: 98 Resp: 585547
Ion Ratio Lower Upper
98 100
100 65.6 50.1 75.1



#62
4-Bromofluorobenzene
Concen: 52.100 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.006 min
Lab File: VY015568.D
Acq: 15 Sep 2023 14:41

Tgt Ion: 95 Resp: 201007
Ion Ratio Lower Upper
95 100
174 84.5 0.0 178.2
176 82.5 0.0 173.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.490 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY015568.D

Acq: 15 Sep 2023 14:41

Instrument :

MSVOA_Y

ClientSampleId :

WC-6(0-6)GRAB





Tgt Ion:117 Resp: 454418

Ion Ratio Lower Upper

117 100

82 54.4 43.4 65.0

119 32.6 26.3 39.5



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.428 min Scan# 1904

Delta R.T. 0.006 min

Lab File: VY015568.D

Acq: 15 Sep 2023 14:41

Tgt Ion:152 Resp: 211315

Ion Ratio Lower Upper

152 100

115 57.6 28.8 86.5

150 156.6 0.0 348.8

