

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091523\
 Data File : VY015574.D
 Acq On : 15 Sep 2023 17:02
 Operator : SY/MD
 Sample : 04399-03
 Misc : 4.47g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11985

Quant Time: Sep 15 17:37:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	299791	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	505527	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	458645	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	215824	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	191418	59.563	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.120%
35) Dibromofluoromethane	7.716	113	159358	54.730	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.460%
50) Toluene-d8	10.179	98	595600	56.485	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	112.960%
62) 4-Bromofluorobenzene	12.477	95	205163	52.411	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	104.820%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091523\
Data File : VY015574.D
Acq On : 15 Sep 2023 17:02
Operator : SY/MD
Sample : 04399-03
Misc : 4.47g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11985

Quant Time: Sep 15 17:37:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 31 20:01:47 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.006 min
Lab File: VY015574.D
Acq: 15 Sep 2023 17:02




Instrument :
MSVOA_Y
ClientSampleId :
72-11985


Tgt Ion:168 Resp: 299791
Ion Ratio Lower Upper
168 100
99 55.3 43.4 65.0



#33
1,2-Dichloroethane-d4
Concen: 59.563 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. -0.000 min
Lab File: VY015574.D
Acq: 15 Sep 2023 17:02

Tgt Ion: 65 Resp: 191418
Ion Ratio Lower Upper
65 100
67 52.4 0.0 101.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.006 min

Lab File: VY015574.D

Acq: 15 Sep 2023 17:02

Instrument :

MSVOA_Y

ClientSampleId :

72-11985





Tgt Ion:114 Resp: 505527

Ion Ratio Lower Upper

114 100

63 19.1 0.0 42.4

88 15.4 0.0 30.6



#35

Dibromofluoromethane

Concen: 54.730 ug/l

RT: 7.716 min Scan# 967

Delta R.T. -0.000 min

Lab File: VY015574.D

Acq: 15 Sep 2023 17:02




Tgt Ion:113 Resp: 159358

Ion Ratio Lower Upper

113 100

111 102.9 81.3 121.9

192 19.9 16.9 25.3








#50
Toluene-d8
Concen: 56.485 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. -0.000 min
Lab File: VY015574.D
Acq: 15 Sep 2023 17:02




Instrument :
MSVOA_Y
ClientSampleId :
72-11985


Tgt Ion: 98 Resp: 595600
Ion Ratio Lower Upper
98 100
100 65.2 50.1 75.1



#62
4-Bromofluorobenzene
Concen: 52.411 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. -0.000 min
Lab File: VY015574.D
Acq: 15 Sep 2023 17:02

Tgt Ion: 95 Resp: 205163
Ion Ratio Lower Upper
95 100
174 85.5 0.0 178.2
176 81.3 0.0 173.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.489 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY015574.D

Acq: 15 Sep 2023 17:02

Instrument :

MSVOA_Y

ClientSampleId :

72-11985





Tgt Ion:117 Resp: 458645

Ion Ratio Lower Upper

117 100

82 53.7 43.4 65.0

119 32.5 26.3 39.5



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.422 min Scan# 1903

Delta R.T. -0.000 min

Lab File: VY015574.D

Acq: 15 Sep 2023 17:02

Tgt Ion:152 Resp: 215824

Ion Ratio Lower Upper

152 100

115 58.5 28.8 86.5

150 157.6 0.0 348.8

