

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091621\
 Data File : VY006046.D
 Acq On : 16 Sep 2021 14:34
 Operator : SY/MD
 Sample : M3733-07
 Misc : 6.96G/5ML/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NAPL-AREA-02-(5-8)

Quant Time: Sep 17 02:45:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 15 05:01:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	156066	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	303136	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	332067	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	171918	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	128288	59.18	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.36%
35) Dibromofluoromethane	7.728	113	95279	50.60	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.20%
50) Toluene-d8	10.185	98	401404	50.02	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.04%
62) 4-Bromofluorobenzene	12.483	95	168705	61.92	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	123.84%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.259	62	3529	1.52	ug/l	94
27) cis-1,2-Dichloroethene	6.996	96	21824	11.20	ug/l	90
64) Tetrachloroethene	10.721	164	10839	3.99	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091621\
Data File : VY006046.D
Acq On : 16 Sep 2021 14:34
Operator : SY/MD
Sample : M3733-07
Misc : 6.96G/5ML/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NAPL-AREA-02-(5-8)

Quant Time: Sep 17 02:45:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 15 05:01:47 2021
Response via : Initial Calibration

