

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091621\
 Data File : VY006052.D
 Acq On : 16 Sep 2021 16:52
 Operator : SY/MD
 Sample : M3757-01
 Misc : 5.08G/5ML/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 AU-06-091521

Quant Time: Sep 17 02:46:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 15 05:01:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	145131	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	272303	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	288743	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	135460	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	109335	54.235	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 108.460%		
35) Dibromofluoromethane	7.722	113	83988	49.658	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 99.320%		
50) Toluene-d8	10.185	98	358977	49.799	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 99.600%		
62) 4-Bromofluorobenzene	12.483	95	137008	55.981	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 111.960%		
Target Compounds						
						Qvalue
16) Acetone	3.954	43	17696	32.004	ug/l	92
20) Methylene Chloride	4.722	84	33110	16.035	ug/l	95
27) cis-1,2-Dichloroethene	6.984	96	2581	1.424	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091621\
Data File : VY006052.D
Acq On : 16 Sep 2021 16:52
Operator : SY/MD
Sample : M3757-01
Misc : 5.08G/5ML/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
AU-06-091521

Quant Time: Sep 17 02:46:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 15 05:01:47 2021
Response via : Initial Calibration

