

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091721\
 Data File : VY006064.D
 Acq On : 17 Sep 2021 12:32
 Operator : SY/MD
 Sample : M3750-02RE
 Misc : 6.80G/5ML/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NAPL-AREA-03-(25-27)RE

Quant Time: Sep 20 01:39:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 15 05:01:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	6844	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	11017	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	10861	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	4597	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	4904	51.585	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 103.160%	
35) Dibromofluoromethane	7.728	113	3542	51.762	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 103.520%	
50) Toluene-d8	10.185	98	13822	47.393	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 94.780%	
62) 4-Bromofluorobenzene	12.483	95	4435	44.790	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 89.580%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.259	62	867	8.541	ug/l #	74
16) Acetone	3.948	43	736	28.226	ug/l #	76
27) cis-1,2-Dichloroethene	6.996	96	1910	22.342	ug/l	87
44) Trichloroethene	8.947	130	472	5.555	ug/l	98
64) Tetrachloroethene	10.727	164	648	7.295	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091721\
Data File : VY006064.D
Acq On : 17 Sep 2021 12:32
Operator : SY/MD
Sample : M3750-02RE
Misc : 6.80G/5ML/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NAPL-AREA-03-(25-27)RE

Quant Time: Sep 20 01:39:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 15 05:01:47 2021
Response via : Initial Calibration

