

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091721\
 Data File : VY006079.D
 Acq On : 17 Sep 2021 18:42
 Operator : SY/MD
 Sample : M3786-01
 Misc : 4.61G/5ML/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CHRT-26878

Quant Time: Sep 20 01:43:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 15 05:01:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	127636	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	241372	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	279475	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	146424	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	99243	55.977	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.960%
35) Dibromofluoromethane	7.728	113	74812	49.901	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.800%
50) Toluene-d8	10.184	98	329516	51.570	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.140%
62) 4-Bromofluorobenzene	12.483	95	137723	63.484	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	126.960%
Target Compounds						Qvalue
18) Methyl Acetate	4.490	43	1385	1.131	ug/l #	78
20) Methylene Chloride	4.722	84	16527	6.885	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091721\
Data File : VY006079.D
Acq On : 17 Sep 2021 18:42
Operator : SY/MD
Sample : M3786-01
Misc : 4.61G/5ML/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CHRT-26878

Quant Time: Sep 20 01:43:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091421S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 15 05:01:47 2021
Response via : Initial Calibration

