

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091724\
 Data File : VY019590.D
 Acq On : 17 Sep 2024 16:59
 Operator : SY/MD
 Sample : P4002-07RE
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-5RE

Quant Time: Sep 18 01:34:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	54221	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	82384	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	52721	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.352	152	12743	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	20589	41.008	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	82.020%
35) Dibromofluoromethane	7.634	113	21994	47.086	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.180%
50) Toluene-d8	10.109	98	89963	52.071	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	104.140%
62) 4-Bromofluorobenzene	12.408	95	19300	29.052	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	58.100%
Target Compounds						
16) Acetone	3.860	43	6232	52.102	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091724\
Data File : VY019590.D
Acq On : 17 Sep 2024 16:59
Operator : SY/MD
Sample : P4002-07RE
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-5RE

Quant Time: Sep 18 01:34:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

