

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091724\
 Data File : VY019592.D
 Acq On : 17 Sep 2024 17:45
 Operator : SY/MD
 Sample : P4002-03RE
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-2RE

Quant Time: Sep 18 01:34:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	101290	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	179240	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	146719	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	49105	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	60269	64.258	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 128.520%	
35) Dibromofluoromethane	7.634	113	57640	56.718	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 113.440%	
50) Toluene-d8	10.109	98	206755	55.004	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 110.000%	
62) 4-Bromofluorobenzene	12.408	95	61493	42.546	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 85.100%	
Target Compounds						
16) Acetone	3.860	43	7521	33.659	ug/l	# 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091724\
Data File : VY019592.D
Acq On : 17 Sep 2024 17:45
Operator : SY/MD
Sample : P4002-03RE
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-2RE

Quant Time: Sep 18 01:34:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

