

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091823\
 Data File : VY015622.D
 Acq On : 18 Sep 2023 18:10
 Operator : SY/MD
 Sample : 04419-19
 Misc : 5.68g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BLIND-DUPLICATE-1

Quant Time: Sep 19 02:05:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 31 20:01:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	140845	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.697	114	239508	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.490	117	218500	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	103542	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	92341	61.160	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 122.320%	
35) Dibromofluoromethane	7.722	113	76813	55.681	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 111.360%	
50) Toluene-d8	10.185	98	282540	56.556	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 113.120%	
62) 4-Bromofluorobenzene	12.483	95	95657	51.578	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 103.160%	
Target Compounds						
17) Carbon Disulfide	4.198	76	6108	1.459	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091823\
Data File : VY015622.D
Acq On : 18 Sep 2023 18:10
Operator : SY/MD
Sample : 04419-19
Misc : 5.68g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BLIND-DUPLICATE-1

Quant Time: Sep 19 02:05:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y082923S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 31 20:01:47 2023
Response via : Initial Calibration

