

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091824\
 Data File : VY019601.D
 Acq On : 18 Sep 2024 13:24
 Operator : SY/MD
 Sample : P4086-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SOIL

Quant Time: Sep 19 01:24:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	324428	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	624868	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	549976	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	204111	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	199680	66.468	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 132.940%	
35) Dibromofluoromethane	7.634	113	197614	55.778	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 111.560%	
50) Toluene-d8	10.109	98	741577	56.590	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 113.180%	
62) 4-Bromofluorobenzene	12.408	95	236025	46.842	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 93.680%	
Target Compounds						
16) Acetone	3.885	43	10423	14.564	ug/l	# 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091824\
Data File : VY019601.D
Acq On : 18 Sep 2024 13:24
Operator : SY/MD
Sample : P4086-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SOIL

Quant Time: Sep 19 01:24:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

