

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091824\
 Data File : VY019602.D
 Acq On : 18 Sep 2024 13:47
 Operator : SY/MD
 Sample : P4090-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-04-091724

Quant Time: Sep 19 01:24:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	323877	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.615	114	617111	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	547978	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	208314	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	191300	63.787	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.580%
35) Dibromofluoromethane	7.634	113	197383	56.413	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.820%
50) Toluene-d8	10.109	98	733643	56.689	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	113.380%
62) 4-Bromofluorobenzene	12.407	95	238651	47.959	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	95.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091824\
Data File : VY019602.D
Acq On : 18 Sep 2024 13:47
Operator : SY/MD
Sample : P4090-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-04-091724

Quant Time: Sep 19 01:24:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

