

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091824\
 Data File : VY019603.D
 Acq On : 18 Sep 2024 14:11
 Operator : SY/MD
 Sample : P4097-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RT2299

Quant Time: Sep 19 01:25:03 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	5449	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.610	114	10401	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	6791	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.359	152	918	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	3915	77.592	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	155.180%
35) Dibromofluoromethane	7.628	113	3002	50.906	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.820%
50) Toluene-d8	10.109	98	11324	51.916	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	103.840%
62) 4-Bromofluorobenzene	12.414	95	1857	22.141	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	44.280%
Target Compounds						
16) Acetone	3.867	43	464	38.601	ug/l	# 36

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091824\
Data File : VY019603.D
Acq On : 18 Sep 2024 14:11
Operator : SY/MD
Sample : P4097-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RT2299

Quant Time: Sep 19 01:25:03 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

