

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091824\
 Data File : VY019609.D
 Acq On : 18 Sep 2024 16:31
 Operator : SY/MD
 Sample : P4087-09
 Misc : 4.87g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NB-614-VOC-02

Quant Time: Sep 19 01:26:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 09 16:00:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	60051	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	93002	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	61962	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	17198	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	22816	41.032	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	82.060%
35) Dibromofluoromethane	7.634	113	25500	48.359	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.720%
50) Toluene-d8	10.103	98	101547	52.065	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	104.140%
62) 4-Bromofluorobenzene	12.402	95	23144	30.861	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	61.720%
Target Compounds						
16) Acetone	3.854	43	2194m	16.562	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091824\
Data File : VY019609.D
Acq On : 18 Sep 2024 16:31
Operator : SY/MD
Sample : P4087-09
Misc : 4.87g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NB-614-VOC-02

Quant Time: Sep 19 01:26:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y090924S.M
Quant Title : SW846 8260
QLast Update : Mon Sep 09 16:00:30 2024
Response via : Initial Calibration

