

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091923\
 Data File : VY015643.D
 Acq On : 19 Sep 2023 14:03
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Sep 20 01:51:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 20 01:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	137429	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	233919	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	211726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	100966	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	87778	58.050	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.100%
35) Dibromofluoromethane	7.716	113	75413	53.863	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.720%
50) Toluene-d8	10.179	98	272352	57.371	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	114.740%
62) 4-Bromofluorobenzene	12.483	95	94321	50.630	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	101.260%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.716	84	39659	19.418	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY091923\
Data File : VY015643.D
Acq On : 19 Sep 2023 14:03
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Sep 20 01:51:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y091923S.M
Quant Title : SW846 8260
QLast Update : Wed Sep 20 01:39:40 2023
Response via : Initial Calibration

